

## Delivery Check List Series 107 2011

Medi-Chair Serial Number # \_\_\_\_\_

### Positioning and Fit

Feature/function checked	Yes	No	NA	Comments
Is back height comfortable?				
Is back width comfortable?				
Is back cushion fit comfortable?				
Is Arm height comfortable?				
With knee bolster cushion adjusted under kneecap, is there ½ to 1-inch gap in the full stand position?				
Are shoe holders adjusted so feet are in front of cut outs in knee bolster?				
Are footrests adjusted so legs are dropped (not up in the front)?				
Looking at the client in the standing position from the front, are the legs straight?  If no, consider changing knee bolster cushion insert to new dimensions. Report new dimensions in comments.				
Does right foot roll over in the standing position?				
Does left foot roll over in the standing position?				
Has excess headrest rod been cut off?				

Use Instructions

Feature/function instructions	Yes	No	NA	Comments
If Roho cushion, did client receive instructions about inflating and deflating cushion?				
If Roho Quadtro cushion, did client receive instructions about how to use the lock and unlock mechanism?				
Did client receive instructions about how to inflate and deflate tires?				
Did client receive instructions on use of chassis anchoring system?				
Was the chassis anchoring bolt and nut both check to make sure they are tight?				
Did chassis anchoring system work in the client's vehicle?				
Did client receive instructions about how to use the circuit breaker?				
Did client receive instructions about how to engage and disengage motor gearbox using levers on gearboxes?				
Did client receive instructions about how to charge batteries and to charge batteries when chair is not in use for periods longer than a week?				
Did client receive instructions on use of the seat belt including safety warning about not letting the seat belt dangle free?				

Did client receive instructions on the use of shoulder harness and/or chest strap including safety warning about the use of upper body restraints anytime when using the stand position?				
Did you go over routine maintenance with client?				
Did client receive instructions about standing tilt and knee bolster pressure?  Did client receive instructions on use of recline and ability to hyperextend?				Client acknowledges instruction _____
Did client have and understand a operator's manual?				
Instruct the client to check for pressure point skin redness after standing, reclining and lifting legs? Call us with any concern.				

### **Chair Acceptance**

Clients  
Signature \_\_\_\_\_ Date \_\_\_\_\_

Representative  
Signature \_\_\_\_\_ Date \_\_\_\_\_

Arbitration: Any dispute concerning the Series 107 Redman Power Chair's performance and/or its factory options' performance or any damages that may be incurred by reason of the Series 107 Redman Power Chair whether they be in tort or in contract, shall be subject to resolution by arbitration.

The arbitration shall be held in accordance with the Uniform Rules of Arbitration and the Federal Arbitration Act. The site of the hearing shall be in Tucson, Arizona. Any witness or party may appear at the hearing via telephone. Any party shall have the right to arbitration: the parties are waiving their right to a trial by jury. A judge or the Superior Court of Pima County Arizona shall appoint the Arbitrator. The Arbitrator shall have the right to order discovery and shall award attorney fees to the prevailing party.

Note: The client understands that only when she/he signs above to indicate their acceptance of the Arbitration clause, that the purchaser is entitled to a 6-month extension to the replacement parts portion of the limited warranty. This 6-month replacement parts extension means that Medi-Chair will provide replacement parts only (no labor or labor related service or travel fees) at no cost to the purchaser for the 6 month period beginning the 13<sup>th</sup> month after the original delivery date and ending at the end of the 18<sup>th</sup> month after the original delivery date.

If the buyer does not initial above to indicate their acceptance of the Arbitration clause above, the buyer is not entitled to an additional 6-month replacement parts only limited warranty on the Series 107 Redman Power Chair.

The buyer also understands that by initialing in the space above, the buyer is giving up certain rights, among those is the right to file a claim in court.